DIAGON
Solutions beyond imagination

INSTRUCTION FOR USE

# Coag Cleaner

## WASHING SOLUTION

**Cat. No.: h60102** **1 x 4,5 L**

### PRODUCT NAME

Coag Cleaner washing solution.

### INTENDED USE

**(For In Vitro Diagnostic Use Only)**

Coag Cleaner solution is the washing solution of DIAGON Coag Line automated coagulation analysers for diagnostic measurements.

### SUMMARY AND EXPLANATIONS

Coag Cleaner solution is a surfactant containing washing solution, that serves for daily maintenance and cleaning of coagulation analysers.

### ACTIVE INGREDIENTS

Coag Cleaner solution contains surfactant (<1,0%) and preservative (<0,5%).

### PRECAUTIONS

- Person installing the Coag Cleaner solution must be a trained laboratory professional!
- Solution coming into contact with specimens and other materials should be handled as if capable of transmitting infection and should be disposed of with proper precautions!
- Avoid microbial contamination of the solution otherwise erroneous results may occur!
- All solutions, waste and utilized disposable laboratory equipment should be considered as hazardous waste! Their handling and disposal should be done according to the valid hazardous material processing regulation.
- Do not use the solution beyond the expiration date printed on the label!

### PREPARATION

Coag Cleaner solution is ready to use. Wait until the solution reaches the working temperature!

### TEST PROCEDURE

For proper application of Coag Cleaner solution, please follow the instructions of the instrument's manual.

### STORAGE AND STABILITY

Coag Cleaner solution in intact cubitainer is stable until the expiration date given on the label, when stored at 15-35°C. Stability after opening in the original cubitainer is shown in below table:

| T (°C) | 15-35 |
|---|---|
| Day | 60 |

### EXPECTED RESULTS

Coag Cleaner solution will effectively clean the tubes in the instrument used on daily basis and keeping carryover values within specification.

### MATERIALS REQUIRED BUT NOT PROVIDED

- Coagulation reagents
- DIAGON Coag Line automated coagulation analyser for measuring.

**DIAGON LTD.**
Baross u. 48-52, 1047 Budapest, Hungary
Tel.: +36 1 3696500
Fax.: +36 1 3696301
Web: www.diagon.com
E-mail: diagon@diagon.com

| SYMBOLS | | | |
|---|---|---|---|
| | Manufacturer | | Use-by date |
| LOT | Batch code | REF | Catalogue number |
| | Keep dry | 35 °C 15 °C | Temperature limit |
| | Handle with care | | Consult instruction for use |
| | Do not use hand hooks | IVD | *In vitro* diagnostic medical device |
| CE | CE mark | | This side up |