Quality Control Certificate

Product Description:	Optitrol Chagas-1
Catalog Number:	SR16027
Lot:	10806998 (Box) 10806996 (Vial)
Expiry Date:	31 Dec 2028
Preservative:	Sodium Azide; ProClin® 950
Application:	For *in vitro* diagnostic use
Storage:	2 to 8 °C
Human Viral Marker:	The controls are prepared from processed human plasma or serum reactive for antibodies against Trypanosoma cruzi IgG. The used human source materials have been tested and found negative for HBsAg and HIV 1+2 and HCV antibodies.
CE 2797:	This product is manufactured and distributed in conformity with the Regulation (EU) 2017/746 of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices.
Batch Release:	The lot has been tested according to valid directions. It meets all required specification criteria and was released for sale.

Approval:

Quality Assurance

19 JUN 2026
Date